Towards *in situ* Backlash Estimation of Continuum Robots using an Endoscopic Camera

Thibault Poignonec, Philippe Zanne, Benoît Rosa and Florent Nageotte¹

Abstract—Accurate control of continuum robots requires handling non-linear behaviors between actuators and distal effectors. In this paper, we develop a method for estimating the non-linearities of tendon-driven degrees of freedom of flexible endoscopic systems by using a distal endoscopic camera and encoders at the proximal side. The proposed approach separates the non-linearities in two parts, namely a pure non uniform backlash and a non-linear function. The backlash is estimated without relying on any model, while the non-linear function is obtained by a pose estimation process. Experiments realized on a robotic flexible endoscopy platform (STRAS) show the validity of the approach for estimating *in situ* the quasi-static behavior of the robot and for compensating the non-linearities of the motion transmission.

Index Terms—Flexible robots. Surgical Robotics: Steerable Catheters/Needles. Computer Vision for Medical Robotics. Tendon/Wire Mechanism.

I. INTRODUCTION

CONTINUUM robots are composed of flexible, slender mechanical structures which, when actuated, can take continuous curvatures. This property makes them suitable for minimally invasive surgery, where access to clinical targets requires using small diameter instruments [1]. Several systems have been designed for surgical endoscopy, where an endoscopic camera provides visual feedback of the instruments [2], [3] (Fig. 1). Accurate control of such continuum robots requires handling complex non-linearities, especially in tendon-driven systems. Several approaches have been used to model these non-linearities [4], [5] and compensate them [6], [7], by relying on a distal measurement. Such measurements can be obtained with embedded sensors [8], [9], which are however not easy to integrate in small diameter instruments. Alternative solutions [10], [11] learn non-linearities offline, assuming that the behavior of the robot does not change during subsequent use. This assumption can be broken by the modification of the shape of the shaft of the endoscope as shown in [12]. Moreover, the relation also changes over time, even for a same endoscope shape (Fig. 2). This advocates for an *in situ* estimation of the non-linearities.

Manuscript received: February, 24, 2020; Revised May, 19, 2020; Accepted June, 14, 2020.

This paper was recommended for publication by Editor Dezhen Song upon evaluation of the Associate Editor and Reviewers' comments. This work was supported by ITMO Cancer under the ROBOT project and by French state funds managed by the ANR within the Investissements d'Avenir program under the ANR-11-LABX-0004 (Labex CAMI). The authors thank Karl Storz for providing the Anubiscope and the flexible instruments.

¹The authors are with ICube, University of Strasbourg, CNRS, France. Nageotte@unistra.fr

Digital Object Identifier (DOI): see top of this page.

Fig. 1. Example view of the considered systems. Here c is the configuration variable representing bending of a bending section of the robot

Non-linear behaviors originate mainly from complex phenomena between the motors actuation and the resulting motion at the distal tip of the continuum robot. For estimating non-linearities *in situ*, at least a sensor measuring the position of the tip and one measuring the proximal position are needed. It is desirable, in endoscopy, to use the camera, which is readily available at the distal part of the endoscope, because integrating sensors in the small diameter instruments is still a practical problem.

The method developed by Reilink et al. [13] allows estimating and compensating backlash by using the endoscopic camera. Corrected motor positions are reconstructed from the images by relying on a model of the continuum robot, which is equivalent to an implicit robot pose estimation. As shown in [14], such models are inaccurate in realistic conditions, due to the uncertain position of the instrument in the endoscope channel, which impacts the eye-to-hand registration. Moreover, the approach only allows for compensating uniform backlash. As shown in Fig. 2, the characteristics of cable driven instruments actually present complex non-linearities. Baek et al. [15] recently proposed compensating hysteresis by fusing image and kinematics information for bending angle estimation and have shown robustness to image occlusions and guide shape variations. The method however relies on the offline training of neural networks and therefore cannot be directly applied *in-situ* for new instruments.

In this work we propose to estimate the backlash by using the embedded camera at the tip of the endoscope, which looks at a moving continuum robot (see Fig. 1). We show that the problem of estimating the global non-linear behavior of the bending DOF of the instrument can be split in two sub-problems : estimating a pure backlash, whose width can

depend on the configuration of the robot, and then estimating the remaining function. We show that the pure backlash can be estimated from image movement detection, without measuring the pose of the robot. This is an interesting feature, as pose estimation from monocular camera in *in vivo* environments is a complex problem. In cases where the pose can be estimated, it is then possible to estimate the remaining function while keeping the initial pure backlash estimate.

The paper is organized as follows. Section II presents the model used in the development of the method. Section III and IV respectively describe the proposed approach for pure backlash and remaining function estimation. Section V presents the validation of the method and results obtained both for predicting the behavior of the robot and compensating the non-linearities. A discussion concludes the paper.

II. MODELING OF ROBOT BEHAVIORS

In this work we consider a cable driven continuum robot observed by an endoscopic camera in an eye-to-hand configuration. This configuration is common in many medical robotic devices developed for endoluminal surgery [2], [16]–[18] or single port laparoscopic surgery [3], [19]. Such devices are usually equipped with sensors at the proximal side only (i.e. encoders mounted onto the drive system). This minimal setup, where no additional sensor –except the endoscopic camera– is available at the distal side, is the case considered in this study (see Fig. 1).

The state of the continuum robot can be represented in different spaces:

- in the actuator space, where it is described by a vector of motor positions $\mathbf{q}$
- in the configuration space, where a vector of configuration variables $\mathbf{c}$ describes the geometric arrangement of the robot
- in the task space, for instance by the position ${}^{base}P_{3D}$ and orientation ${}^{base}R_{eff}$ of the continuum robot effector with respect to the base frame
- in the image space, by the perspective projection of points of the continuum robot body, for instance Q , the projection of P_{3D} .

For each space, the parameters to be used and their number required for representing the robot depend on its design.

In this work, we assume that the relation between the configuration space and the task space is known, *i.e.* one can compute the position/orientation of the robot end effector given a set of configuration variables. This can be done for instance by relying on a geometric model of the robot (for instance a constant curvature model [20]). It is written as: $P_{3D} = g(\mathbf{c}), g: \mathbb{R}^n \rightarrow \mathbb{R}^3$.

On the contrary, the link between the actuator space and the configuration space $\mathbf{c} = \mathbf{f}(\mathbf{q})$ is only coarsely known. In the considered systems, it is usual that conventional kinematic models do not account for complex behaviors in the motion transmission. This is especially true for robots with flexible shafts and for the tendon-driven DOFs. Some of the observed behaviors are shown in Fig. 2. A coarse function denoted f_{coarse} is assumed to be available, for instance a simple

Fig. 2. Non-linear behaviors measured on a tendon driven DOF of a continuum robot, between proximal actuator and distal configuration. The different colors correspond to different dates of measurement between *in vivo* experiments. Strong variations can be observed.

kinematic model assuming perfect transmission. The core of this work is to estimate the actual relation between actuator space and configuration space.

The transformation between the task space and the image space is also only coarsely known. The camera is assumed calibrated and a nominal eye-to-hand transform ${}^{cam}t_{base}, {}^{cam}R_{base}$ is available, which allows estimating $Q \sim \hat{Q} = h(P_{3D})$. However, as shown in [14], the hand-eye calibration in endoscopic systems includes uncertainties linked to the necessary mechanical play for translating and rotating instruments in the working channel.

A. Modeling of the quasi-static non-linearities

We consider one tendon-driven degree of freedom of the continuum robot. The non-linearity in the behavior of this DOF can be represented by the hysteresis shape relating q (one element of $\mathbf{q}$) to the obtained configuration variable c (one element of $\mathbf{c}$), usually a bending angle. As shown in Fig. 2 the hysteresis can have a complex shape. In our approach we propose to decompose this global non-linearity in two parts, as shown in Fig. 3:

1) A backlash, which appears at a change of direction of the actuator. This is responsible for the hysteresis shape of the characteristic curve. In the following we call this part “pure backlash”. The approach proposed in [13] considered a constant backlash. There is, however, no reason to consider the backlash constant *a priori*. Therefore our method identifies the relation between the actuator position and the width of the backlash and express it as a function $B(q)$ (see section III). This modeling implies that the backlash is symmetric, *i.e.* for a given distal configuration the absence of motion happens on the same motor range, whatever the direction of the motor motion.

2) A function $f_{remain}(q)$, generally non-linear, which links the actuator position to the configuration variable value once pure backlash has been removed or compensated. In the following we call this function the remaining function. This function can also be viewed as the central line of the hysteresis

Fig. 3. Proposed modeling of the continuum robot, where transmission non-linearities are split in two parts.

Fig. 4. General scheme used for non-linearities estimation. The upper part of the block-scheme (blocks 1 to 3) concerns estimation of pure backlash while the lower part (blocks 4,5) processes the remaining function. They are combined to provide the complete estimate (block 6).

curve. It should be noted that it is generally different from the envelope functions of the hysteresis (see for instance Fig. 6).

Representing complex hysteresis shapes by combining a backlash operator and a function, as shown in Fig. 3, has for instance been proposed in [21]. However, identifying the two parts separately is a novel approach. It is of interest here because the two parts can be obtained from different information as will be shown in the following. The next sections present the proposed method in details. The general scheme is depicted in Fig. 4.

III. PURE BACKLASH ESTIMATION

A. Estimation of the backlash at a given motor position

Pure backlash can be characterized by the absence of end-effector motion after a change of direction of the actuator at the proximal side. The backlash width is then simply the range of motor motion which does not create distal motion. The proposed approach therefore consists in detecting the appearance of motion after a change of motor direction. The absence of motion and the appearance of motion are events which are directly visible in the image and which do not require a 3D pose estimation process.

The process for backlash estimation is as follows (Fig. 5):

- The position of the effector is tracked in the image
- An actuator direction change (change of $\text{sign}(\dot{q})$) is requested at time t_1 and the corresponding motor position q_1 is recorded
- When a movement is detected in the image at time t_2 , the corresponding position q_2 is recorded

The backlash width can be considered as a function of the distal configuration c . However, c is not directly accessible with the considered setup and its estimation would require a pose estimation process. Therefore, it is more convenient to describe the backlash width as a function of an actuator position, which is readily available. We have chosen to express the backlash width as a function of the mean actuator position

Fig. 5. Pure backlash estimation scheme. Left: actuator motions for a requested change of direction, and the detected motion in the image in function of time. Top-right: corresponding evolution in the configuration vs actuator space. Bottom-right: The measured backlash width provides a point in the backlash width function.

in the backlash zone $q_{\text{mean}} = \frac{q_1 + q_2}{2}$. The value of width is then given by $B(q_{\text{mean}}) = |q_2 - q_1|$ and it can represent backlash for both direction changes (positive to negative velocity and negative to positive velocity). Actually, since different sources of information are used to obtain t_1 and t_2 a correction is needed to temporally align image measurements and motor measurements. This is especially important when using low frame rate image acquisition systems, which also introduce delays. We then use

$$B(q_{\text{mean}}) = |q(t_2 - t_{\text{delay}}) - q_1|, \quad (1)$$

where t_{delay} is the temporal delay introduced by the image acquisition and processing pipeline. It is assumed here that the motor positions are obtained at a high frame rate with negligible delay as is usually the case.

This simple process is independent of any robot model or 3D position estimation process. It is therefore also independent on the calibration parameters of the camera and of the eye-to-hand transform ${}^{\text{cam}}T_{\text{base}}$.

B. Construction of the backlash width function

As the backlash width depends on q , the method presented in section III-A must be performed at different actuators positions, in order to estimate the function $B(q)$ over a useful range of actuators positions. For an *in situ* estimation prior to use, typically after the medical robot insertion in the body but before the actual surgical use, a pre-programmed motor trajectory can be used to reconstruct the backlash width function B . This trajectory can consist in several small range back and forth motions superimposed on a ramp which covers the full range of the useful motor positions. Using this pre-programmed trajectory, one can estimate values of $B(q)$ over the desired range, but only at discrete intervals. Using the backlash model for compensation (see section III-C) requires however a continuous function for all motor positions. We propose to interpolate and filter the obtained values by using cubic B-splines, with N control points homogeneously spread over the motor range. This choice was made based on the observation that hystereses obtained with external sensors in [10] exhibit an overall smooth variation but with possible high local slopes.

Fig. 6. Reconstructed hystereses, by combining a linear f_{coarse} function with pure backlash (top) and by combining estimated f_{remain} with backlash (bottom).

Given the initial coarse model of the $q \rightarrow c$ mapping $c = f_{coarse}(q)$ the direct complete model of the non-linearities can be obtained by combining the backlash model and the f_{coarse} function. The external envelope defined by functions f_1 and f_2 can be obtained by a numerical reindexing of f_{coarse} :

$$f_{1,2}(q_{off}) = f_{coarse}(q) \text{ with } q_{off} = q \mp \frac{B(q)}{2} \quad (2)$$

This is illustrated in Fig. 6 (top), where f_{coarse} is linear. Note that because B generally depends on q , $f_{1,2}$ are not linear. The variable backlash model enforces a deformation of the external envelope.

At a time index k the direct model can be numerically simulated by (see [22]):

$$c[k] = \begin{cases} f_2(q[k]) & \text{if } q[k] > q[k-1] \text{ and } c[k] \geq f_2(q[k]) \\ f_1(q[k]) & \text{if } q[k] < q[k-1] \text{ and } c[k] \leq f_1(q[k]) \\ c[k-1] & \text{otherwise} \end{cases} \quad (3)$$

C. Use for backlash compensation

The identified model can be used to compensate the backlash. Given a reference trajectory $c^*(k)$ for the configuration variable, the compensated reference motor trajectory can be computed as :

$$q^*[k] = \begin{cases} F_{1inv}(c^*[k]) & \text{if } c[k]^* < c[k-1]^* \\ F_{2inv}(c^*[k]) & \text{if } c[k]^* > c[k-1]^* \\ q^*[k-1] & \text{if } c^*[k] = c^*[k-1] \end{cases} \quad (4)$$

where

$$\begin{cases} F_{1inv}(c^*) = \hat{f}_{coarse}^{-1}(c^*) - \frac{B(\hat{f}_{coarse}^{-1}(c^*))}{2} \\ F_{2inv}(c^*) = \hat{f}_{coarse}^{-1}(c^*) + \frac{B(\hat{f}_{coarse}^{-1}(c^*))}{2} \end{cases} \quad (5)$$

$\hat{f}_{coarse}^{-1}$ is the inverse function of f_{coarse} , which can be obtained by a numerical inversion. Smoothing can also be added to transitions to avoid actuator saturation [12].

This compensation can advantageously be used in a telemanipulation framework, for correcting the behavior of the robot when a change of direction of the distal configuration variable is required. It can thus avoid apparent delays in executing the user's reference trajectories.

IV. REMAINING FUNCTION RECONSTRUCTION

For automatic motions realization, as described in [10], the compensation of backlash alone is not sufficient for providing a good positioning accuracy. Indeed, f_{coarse} is usually inaccurate, and therefore $f_{1,2}$ are also. There are for instance dead zones near the straight configuration, as shown in Fig. 2 (green curve, between 0° and 10° for instance), which are difficult to predict from models. This means that even if backlash may be correctly compensated when changes of direction are required, the link between the rate of variation of c^* and of q^* can be incorrect. In other words, the task space position can be inaccurate. For improving accuracy, one can estimate the complete hysteresis envelope, i.e. f_1 and f_2 , from data. These functions link the actuator positions to the configuration variable values as long as no change of direction is applied. This information can be obtained from the endoscopic camera only if the configuration variable can be extracted from the image. This in turn requires a geometric model of the continuum robot, a model of the camera and of the eye-to-hand configuration. Such techniques have been used to reconstruct three [23] or six configuration variables [14]. In the considered case where a single DOF has to be estimated, a marker at the tip of the instrument coupled with a conventional geometric model (for instance a constant curvature model) is sufficient. We propose an approach similar to [23], but instead of reconstructing a corrected motor position, we focus on the configuration variable c . Given a model of the endoscopic camera and of the eye-to-hand calibration, we have :

$$\hat{Q} = h(g(c)) \quad (6)$$

We denote the interaction matrix (image jacobian) which relates the variation of c to the variation of $\hat{Q}$ by J :

$$\frac{d\hat{Q}}{dt} = J \frac{dc}{dt} \text{ with } J = \frac{\partial h \circ g}{\partial c} \quad (7)$$

Given a position of the marker in the image Q , one can estimate the optimal value of c by minimizing the cost function $L = \|Q - \hat{Q}\|^2$ by using an iterative optimization process such as Levenberg-Marquardt:

$$\hat{c}_j = \hat{c}_{j-1} + (J^T J + \lambda(\text{diag}(J^T J))^{-1})^{-1} J^T (Q - \hat{Q}) \quad (8)$$

Note that this approach, as the ones used in [14], [23], minimizes the error of a projection in the image. This scheme can therefore succeed and provide very low projection errors even when models are erroneous. In such cases, the obtained values for the configuration variables will be inaccurate.

By applying this estimation process during the programmed motion of the instrument used for pure backlash estimation as described in section III-B, one can reconstruct the hysteresis characteristic curves (see Fig. 9). After this step, the functions defining the external envelope of the hysteresis could be used to compute the width of the backlash in function of q_{mean} as

for instance in [10]. However, in this case the pure backlash estimation obtained previously would be overwritten. This is not desirable, because contrary to [10] where accurate external sensors are used, the process used here to estimate c is subject to modeling and measurements errors. We therefore propose to keep the initial pure backlash estimation, which was obtained independently of any pose estimation, and to estimate the remaining function. For this purpose the remaining function is estimated as the central fiber of the characteristics, *i.e.* the horizontal median line of the envelope. For extracting it, the characteristic is processed as an image by applying morphological operations for filling holes, extremal horizontal values detection and spline fitting. This provides function $f_{\text{remain}} : q \rightarrow c$, as shown in Fig. 9.

The complete obtained model can for instance be used for an automatic positioning task as will be shown in section V. Given a desired trajectory $c^*[k]$ for the configuration variable, the required actuator position can be obtained using eq. (4), but replacing f_{coarse} with f_{remain}

$$\begin{cases} F_{1\text{inv}}(c^*) = f_{\text{remain}}^{-1}(c^*) + \frac{B(f_{\text{remain}}^{-1}(c^*))}{2} \\ F_{2\text{inv}}(c^*) = f_{\text{remain}}^{-1}(c^*) - \frac{B(f_{\text{remain}}^{-1}(c^*))}{2} \end{cases} \quad (9)$$

V. SIMULATION AND EXPERIMENTS VALIDATION

Several simulations and experiments were carried out to show the validity of the approach and its possible interest.

A. Experimental setup

We consider the left instrument of the STRAS robotic platform [18] (see Fig. 12). The instrument has three DOF : translation, rotation and bending. The bending DOF, which is actuated by antagonist cables is the main focus of the experiments. The robot DOFs are controlled in position by a real-time control PC (see [18] for details). Endoscopic images are acquired at 25 Hz through a framegrabber. Image processing codes are in C++ on a separate computer, which sends high-level commands to the real-time control PC, where inverse backlash models are implemented.

B. Simulation

The proposed approach is first studied in simulation of the above-mentioned system, in order to show the interest of our method in controlled conditions. We simulate the robot using a perfect constant curvature model, with a hysteresis of constant width 15° on the bending DOF. We consider in addition that the instrument can slightly move in its working channel (as shown in [14]). Therefore, we introduce a $\pm 1.5^\circ$ error around the y axis on the hand-eye registration matrix, the sign depending on the distal motion direction. We then reconstruct the pure backlash and remaining function as described in sections III-C and IV, with bending velocities at $15^\circ/\text{s}$. For comparison purposes, we also reconstruct the hysteresis in a single step by directly estimating the configuration variable (see section IV).

Figure 7 shows the obtained hysteresis cycles and the estimated backlash in both cases. As could be foreseen backlash

Fig. 7. Reconstructed hysteresis (top) and backlash width function (bottom) as a function of q . The robot is simulated with a constant 15° backlash on the bending joint, and a $\pm 1.5^\circ$ error on the hand-eye calibration. Green: ground truth, Blue: Direct single-step estimation of the hysteresis using pose estimation, Red: using the proposed approach.

reconstruction is impacted by the slight registration error, even with a perfectly known constant curvature model. Our proposed two-step approach accurately estimates the backlash (mean error 0.7° (4.7%), std. 0.09°), while the model-based estimation features large errors (mean error 6.6° (44%), std. 0.51°).

It has to be noted that this simulation example was chosen on purpose to exhibit limitations of more conventional approaches relying on pose estimation and configuration variable reconstruction. However, it shows that avoiding pose estimation can be useful in realistic cases.

C. Image measurements

The continuum robot end effector needs to be tracked in the image. Since a single DOF is considered, it is sufficient to track a single point. Since the focus of the present work is not on the image processing aspects, we used a simple colored marker attached at the tip of the instrument, which can be efficiently tracked using simple image processing techniques. More advanced marker-based or marker-less methods [14], [15], [23]–[25] can be used for in vivo settings.

D. Backlash estimation

The delay of the image acquisition chain has been estimated to 2 frames (*i.e.* $t_{\text{delay}} = 80\text{ms}$), with an uncertainty $\Delta t_{\text{delay}} = 20\text{ms}$ (half period). Assuming a constant velocity of the actuator v during backlash crossing, this creates an estimation error $\Delta q = \|\Delta t_{\text{delay}} v\|$ on $q(t_2 - t_{\text{delay}})$. It is therefore recommended to apply a low velocity for backlash estimation. Practically, the velocity is set to 5% of the maximum motor velocity, which corresponds approximately to a maximum bending speed of $20^\circ/\text{s}$ at the distal end. The corresponding uncertainty onto the actuator position is 0.025mm of cable displacement. For separating motion and absence of motion, the image position of the marker Q is differentiated and a threshold T is applied. T was set to 0.2 pixels/frame after practically assessing that the noise for still marker localization was less than 0.15 pixels.

Figure 8 shows the backlash estimation as a function of the motor position. The continuous function has been obtained

Fig. 8. Backlash width (measurement (dots) and B-spline based reconstructed shape (red line)) as a function of the mean actuator position. Blue lines show the spline knots positions. Units are in mm of cable displacement.

Fig. 9. Characteristics between the actuator and configuration variable obtained using configuration variable estimation from images.

by B-spline fitting with $N = 14$ knots (8 central and 3 at each extremity of the motor range). One clearly observes the strong increase of the backlash near the center of the motor range (straight configuration of the instrument). This behavior was already observed in [10], [26]. This clearly confirms the need of a variable backlash estimation method (see section II-A).

E. Remaining function estimation

The remaining function was estimated using the same input images and motor data as for the backlash width. The typical actuator / configuration characteristics are shown in Fig. 9. f_{remain} is then extracted as described in section IV.

F. Assessment of prediction capability

For assessing the validity of the pure backlash and the remaining function estimation, a testing motion is used, which consists in periodic trapezoidal back and forth motions realized with the motor actuating bending superposed on a linearly varying mean value (see Fig. 10 and accompanying video). This trajectory includes many changes of directions, in order to exhibit the effects of backlash. The position of the instrument is tracked in the endoscopic image, and the actual trajectory is compared with the modeling obtained with f_{coarse} (pure kinematic model of the instrument [27]), with f_{coarse} combined with backlash estimation only ($f_{coarse} + B$) and with the complete estimation ($f_{remain} + B$) (see Fig. 11). Results are presented in table I.

Model used	Error (in pix.)			
		On Qx	On Qy	distance
f_{coarse}	mean $\pm$ std	78 ± 73	30 ± 5	89 ± 67
	max.	309	39	309
$f_{coarse} + B$	mean $\pm$ std	59 ± 61	31 ± 4	73 ± 53
	max.	289	36	289
$f_{remain} + B$	mean $\pm$ std	17 ± 15	34 ± 1	40 ± 9
	max.	78	40	85

TABLE I

COMPARISON OF PREDICTION ERRORS IN THE ENDOSCOPIC IMAGE USING f_{coarse} , $f_{coarse} + B$ AND $f_{remain} + B$. THE ERRORS ARE SHOWN FOR THE x AXIS (Qx), THE y AXIS (Qy) AND IN NORM (DISTANCE).

When using f_{coarse} only, the norm of the error in the image shows large pseudo-periodic variations, with a mean distance error of 89 pixels. This is mainly due to the prediction going on either side of the actual position, because the absence of movement at the distal tip at the change of direction is not predicted. The combination of the backlash with f_{coarse} allows to greatly decrease the amplitude of the squares because, at each change of direction, the model predicts the range of motor motion that does not create distal motion. Zoom A on Fig. 10 shows that the appearance of motion in the image at each cycle is well synchronized between prediction and measurement. However, the prediction shows larger amplitudes of motions than the measurement, and the mean distance error only decrease to 73 pixels. This arises from an incorrect prediction of the rate of variation at the distal side. Namely, the rate is overestimated because f_{coarse} does not correspond to the actual remaining function. This is especially visible between 150s and 175s (see zoom B) because the instrument comes close to the straight configuration, where the actual rate of motion is very low and the discrepancy with f_{coarse} particularly large. Note, however, that the synchronization of motions remains good, indicating that the backlash is well estimated. When using the complete model with f_{remain} , the velocity of distal movement is better predicted, which allows an important reduction of the prediction error to 40 pixels. In zoom B, the remaining function only allows to slightly decrease the error on the image position prediction. The remaining error can be due to the transition movement between backlash and the envelope function or to coupling effects.

It can be noted that the norm of the error never goes under 40 pixels. This comes from errors in the registration between the endoscopic camera and the robot, which mainly affect the vertical position (Qy) of the instrument in the image. In the presented experiments, the motion of the instrument is mainly along the horizontal axis of the image, so that the error on Qy is almost constant (see table I). The prediction error along the horizontal axis (Qx) is therefore significantly reduced from 78 pixels to 17 pixels when using $f_{remain} + B$.

G. Assessment of compensation capability

The inverse non-linear models have been implemented onto the robotic system. For assessing the effect of compensation, a 2D trajectory (an ellipse) is defined in the task space of the continuum robot. The configuration trajectory (bending, rotation, translation) is computed using the inverse geometric

Fig. 10. Pseudo-periodic trajectory used for testing prediction capability of different models: f_{coarse} (black), $f_{coarse} + B$ (green) and $f_{remain} + B$ (yellow). Zooms A and B show that when backlash is modeled the appearance of motion is well synchronized with the measurement.

Fig. 11. Reprojection errors between the measured position of the effector in the endoscopic image and the prediction of effector position provided by f_{coarse} (black), $f_{coarse} + B$ (green) and $f_{remain} + B$ (yellow).

model of the instrument [27]. The obtained trajectory is then performed in open-loop and observed by an external camera parallel to the plane of the ellipse. Figure 12 provides the qualitative results. The non-linearity compensation allows to largely improve the accuracy of the trajectory.

VI. DISCUSSION AND CONCLUSION

We have presented a method for estimating non-linearities of cable-driven continuum robots, which uses an endoscopic camera conventionally available in endoluminal digestive surgery and which is thus suitable for *in situ* use. The method has been shown valid for free moving continuum instruments through laboratory experiments, both for predicting the robot behavior and for compensating the non-linearities.

One originality of the approach is to separate hysteresis estimation in two parts, and to estimate pure backlash without relying on pose estimation. This is an interesting feature because pose estimation in *in vivo* environments is a complex task [14]. Here, backlash is obtained from motor measurements and image measurements only. Tracking a single point

Fig. 12. Open loop realization of a reference trajectory (yellow), by relying on f_{coarse}^{-1} (blue) and on the inverse model constructed from $f_{remain} + B$ (see eq. 9) (green)

on the instrument in the image is sufficient, and there is no need to combine several image information to estimate a pose or a configuration variable (as for instance proposed in [13], [14], [23]) or to use data acquired beforehand on the system as proposed in [15]. Moreover, methods based on models need to handle singularities in the process of 3D information reconstruction. Such singularities can come from the kinematics of the continuum robot (for instance, rotation is ill-defined around the straight configuration of the instrument) or from the image jacobian which links instrument tip motion to its apparent motion in the image. There are solutions to avoid such singularities, but a specific technique should be designed, which may be difficult to implement in cases where the available models are erroneous. Therefore, the proposed approach is well adapted for estimating backlash *in situ*.

The method also allows to take into account non uniform backlash, which is generally the case as shown in Fig. 2 and 8. When identified, pure backlash can be compensated, which can improve user experience during teleoperation. If pose estimation can be obtained, it is possible to estimate proximal to distal position relation. Interestingly, the proposed approach allows complementing the backlash model, without changing it. Therefore, if pose estimation turns out erroneous because of modeling or measurement errors, the backlash compensation won't be affected. This is an important improvement over the method in [13], where complete hysteresis estimation relies on pose estimation. Indeed, simulations show that slight errors on the camera / robot registration can have a large impact on the estimation of backlash when relying on configuration variable reconstruction (see Fig. 7).

As seen in Fig. 10, 11 and 12, errors remain, which have several identified sources:

1) As expected, the estimation of f_{remain} depends on 3D position estimation, which is subject to modeling errors (non uniform curvature of the robot, for instance).

2) By separating pure backlash and remaining non-linear function, it is assumed that the system directly switches from the pure backlash to the envelope function of the hysteresis. However, it can be observed in Fig. 2 that smooth transitions can exist. These transitions cannot be directly modeled with the proposed approach because they would require reconstruct-

ing a rate of change of the configuration variable.

3) It is also assumed that backlash is symmetric. Complex coupling effects in the motion transmission of the instrument could create non symmetric backlash. This could be one source of the errors observed in Fig. 10, zoom B.

4) The proposed approach mainly deals with quasi static non-linearities. With the given setup, which relies on endoscopic images, it won't be possible to measure dynamic effects such as stick-slip, because the acquisition framerate is too low. In an extension of the work, it could be possible to try decreasing the effects of such dynamic non-linearities, for instance by designing acceleration trajectories that limit their appearance.

It should however be noted that these difficulties also affect other techniques for hysteresis compensation relying on the use of endoscopic images, and as shown in section V the obtained results are satisfactory.

The motions of the guide of the instruments (the endoscope) are also impacted by its own non-linearities. While our work mainly aimed at improving the control of the instruments, which is of higher importance during surgical stages, it could also be useful to compensate non-linearities of the guide during navigation stages. The proposed approach requires to see the motions of the continuum robot in images in an eye-to-hand configuration, and therefore cannot be directly applied to the eye-in-hand configuration of the endoscope. The proposed approach could possibly be extended to estimate motion / absence of motion in such a configuration, for instance using motion-compensated SLAM algorithms [28]. This is however out of the scope of the current article.

Non-linearity estimation, prediction and compensation were realized for free motions of the robot. Contacts would impact the validity of the method. Therefore, it is mainly aimed at non contact tasks, such as optical coherence tomography biopsy or laser dissection. As a future work, we aim at adapting the proposed method to make it "on-line", *i.e.* working during the actual use of the system. A first approach would be to generate very small motions, just larger than the actual backlash, when instruments are not moved by the user. Another approach would be to rely on the motions generated by the user.

REFERENCES

- [1] J. Burgner-Kahrs, D.C. Rucker, and H. Choset, Continuum robots for medical applications: a survey, *IEEE Trans. Robot.*, vol. 31, no.6, pp.1261-1280, Dec. 2015.
- [2] <https://medrobotics.com/gateway/flex-robotic-system/?c=INTL>, commercial website of Medrobotics, accessed on 05/19/20.
- [3] <https://titanmedicalinc.com/technology/>, commercial website of Titan Medical, accessed on 05/19/20.
- [4] V. Agrawal, W. Peine and B. Yao, Modeling of Transmission Characteristics Across a Cable-Conduit System, *IEEE Trans. Robot.*, vol. 26, no. 5, pp. 914-924, Oct. 2010.
- [5] J. Jung et al., A modeling approach for continuum robotic manipulators: Effects of nonlinear internal device friction, in *IEEE/RSJ Int. Conf. Intell. Robots and Systems*, San Francisco, California, Sep. 2011, pp. 5139-5146.
- [6] T. N. Do et al., Hysteresis modeling and position control of tendon-sheath mechanism in flexible endoscopic systems, *Mechatronics*, vol. 24, no.1, pp. 12-22, Feb. 2014.
- [7] Z. Sun, Z. Wang, S.J. Phee, Elongation Modeling and Compensation for the Flexible Tendon-Sheath System, *IEEE/ASME Trans. Mechatronics*, vol. 19, no. 4, pp. 1243-1250, Aug. 2014.
- [8] R. J. Roesthuis, S. Janssen, S. Misra, On using an array of fiber Bragg grating sensors for closed-loop control of flexible minimally invasive surgical instruments, in *IEEE/RSJ Int. Conf. Intell. Robots and Systems*, Tokyo, Nov. 2013, pp. 2545-2551.
- [9] H. Liu et al., Large deflection shape sensing of a continuum manipulator for minimally-invasive surgery, in *Proc. 2015 IEEE Int. Conf. Robot. Automat.*, pp. 201-206.
- [10] R. Aleluia Porto, F. Nageotte, P. Zanne and M. de Mathelin, Position control of medical cable-driven flexible instruments by combining machine learning and kinematic analysis, in *IEEE Int. Conf. Robot. Automat.*, Montreal, May 2019, pp. 7913-7919.
- [11] W. Xu et al., Data-driven methods towards learning the highly nonlinear inverse kinematics of tendon-driven surgical manipulators, *Int. J. Med. Robot. Comput. Assisted Surgery*, vol. 13, no. 3, 2017.
- [12] L. Ott, F. Nageotte, P. Zanne, M. de Mathelin, Robotic Assistance to flexible endoscopy by physiological motion tracking, *IEEE Trans. Robot.*, Volume 27, no 2, pp. 346-359, Jan. 2011.
- [13] R. Reilink, S. Stramigioli and S. Misra, Image-based hysteresis reduction for the control of flexible endoscopic instruments, *Mechatronics*, vol. 23, no.6, pp. 652-658, Sept. 2013.
- [14] P. Cabras et al., An adaptive and fully automatic method for estimating the 3D position of bendable instruments using endoscopic images, *Int. J. Med. Robot. Comput. Assisted Surgery*, vol. 13, no. 4, Dec. 2017.
- [15] D. Baek et al., Hysteresis Compensator with Learning-Based Hybrid Joint Angle Estimation for Flexible Surgery Robots, *IEEE Robot. Automat. Lett.*, Feb. 2020.
- [16] <http://easyendosurgical.com/>, website of EasyEndo, accessed on 05/19/20.
- [17] D. J. Abbott, C. Becke, R. I. Rothstein, and W. J. Peine, Design of an endoluminal notes robotic system, in *IEEE/RSJ Int. Conf. Intell. Robots and Systems*, San Diego, California, Oct. 2007, pp. 410-416.
- [18] L. Zorn et al., A Novel Telemanipulated Robotic Assistant for Surgical Endoscopy: Preclinical Application to ESD, *IEEE Trans. Biomed. Eng.*, vol. 65, no. 4, pp. 797-808, Apr. 2018.
- [19] J. Ding, R. E. Goldman, K. Xu, P. K. Allen, Fowler, and N. Simaan, Design and coordination kinematics of an insertable robotic effectors platform for single-port access surgery, *IEEE/ASME Trans. Mechatronics*, vol. 18, no.5, pp. 1612-1624, Oct. 2013.
- [20] R. J. Webster and B. A. Jones, Design and kinematic modeling of constant curvature continuum robots: A review, *Int. J. Robot. Res.*, vol. 29, no. 13, pp. 1661-1683, Jun. 2010.
- [21] Y. Rochdi et al., Identification of block-oriented systems in the presence of nonparametric input nonlinearities of switch and backlash types, *Automatica*, vol. 46, no.5, pp. 864-877, May 2010.
- [22] J. Vörös, Modeling and identification of systems with backlash, *Automatica*, vol. 46, no. 2, pp. 369-374, Feb. 2010.
- [23] R. Reilink, S. Stramigioli and S. Misra, 3D position estimation of flexible instruments: marker-less and marker-based methods, *Int. J. Comput. Assisted Radiol. and Surgery*, vol. 8, no. 3, pp. 407-417, May 2013.
- [24] A. Reiter, R.E. Goldman, A. Bajo et al. A learning algorithm for visual pose estimation of continuum robots, in *IEEE/RSJ Int. Conf. Intell. Robots and Systems*, San Francisco, California, 2011, pp. 2390-2396.
- [25] B. Rosa, V. Bordoux, F. Nageotte, Combining differential kinematics and optical flow for automatic labelling of continuum robots in minimally invasive surgery, *Frontiers Robot. AI*, Frontiers Research Foundation, Online, Sep. 2019.
- [26] B. Bardou et al., Improvements in the control of a flexible endoscopic system, in *IEEE Int. Conf. Robot. Automat.*, Saint Paul, Minnesota, 2012, pp. 3725-3732.
- [27] A. De Donno et al., Master / Slave Control of Flexible Instruments for Minimally Invasive Surgery, in *IEEE/RSJ Int. Conf. Intell. Robots and Systems*, Tokyo, Nov. 2013, pp. 483-489.
- [28] P. Mountney and G.Z. Yang, Motion compensated SLAM for image guided surgery, in *MICCAI*, Beijing, Sep. 2010, pp. 496-504.