MSCT versus CBCT: evaluation of high-resolution acquisition modes for dento-maxillary and skull-base imaging

Jean-Philippe Dillenseger · Jean-François Matern · Catherine-Isabelle Gros ·
Fabien Bornert · Christian Goetz · Jean-Marie Le Minor ·
André Constantinesco · Philippe Choquet

Received: 23 July 2014 / Revised: 19 August 2014 / Accepted: 8 September 2014
© European Society of Radiology 2014

Abstract

Objectives Our aim was to conduct a quantitative and qualitative evaluation of high-resolution skull-bone imaging for dentistry and otolaryngology using different architectures of recent X-ray computed tomography systems.

Material and methods Three multi-slice computed tomography (MSCT) systems and one Cone-beam computed tomography (CBCT) system were used in this study. All apparatuses were tested with installed acquisition modes and proprietary reconstruction software enabling high-resolution bone imaging. Quantitative analyses were performed with small fields of view with the preclinical vmCT phantom, which permits to measure spatial resolution, geometrical accuracy, linearity and homogeneity. Ten operators performed visual qualitative

analyses on the vmCT phantom images, and on dry human skull images.

Results Quantitative analysis showed no significant differences between protocols in terms of linearity and geometric accuracy. All MSCT systems present a better homogeneity than the CBCT. Both quantitative and visual analyses demonstrate that CBCT acquisitions are not better than the collimated helical MSCT mode.

Conclusion Our results demonstrate that current high-resolution MSCT protocols could exceed the performance of a previous generation CBCT system for spatial resolution and image homogeneity.

Key Points

- *Quantitative evaluation is a prerequisite for comparison of imaging equipment.*
- *Bone imaging quality could be objectively assessed with a phantom and dry skull.*
- *The current MSCT shows better image quality than a dental CBCT system.*
- *CBCT remains a work-in-progress technology.*

J.-P. Dillenseger · J.-F. Matern · C. Goetz · A. Constantinesco ·
P. Choquet
Imagerie Préclinique-UF6237, Pôle d'imagerie, Hôpitaux
Universitaires de Strasbourg, Strasbourg, France

J.-P. Dillenseger · C. Goetz · J.-M. Le Minor · P. Choquet
Icube, équipe MMB, CNRS, Université de Strasbourg, Strasbourg,
France

J.-P. Dillenseger · J.-F. Matern · C.-I. Gros · F. Bornert · C. Goetz ·
J.-M. Le Minor · P. Choquet
Fédération de Médecine Translationnelle de Strasbourg, Faculté de
Médecine, Université de Strasbourg, Strasbourg, France

C.-I. Gros · F. Bornert
Faculté de Chirurgie Dentaire, Université de Strasbourg, Strasbourg,
France

J.-M. Le Minor
Institut d'Anatomie Normale, Université de Strasbourg, Strasbourg,
France

P. Choquet (✉)
Imagerie Préclinique, Biophysique et Médecine Nucléaire, Hôpital
de Hautepierre, 1 Av. Molière, 67098 Strasbourg Cedex, France
e-mail: pchoquet@unistra.fr

Keywords Cone-beam computed tomography · Multi-slice
computed tomography · Bone · Quantitative evaluation ·
Radiological phantoms

Abbreviations

CBCT	Cone-beam computed tomography
CT	Computed tomography
FBP	Filtered back projection
FOV	Field of view
FPD	Flat panel detector
MSCT	Multi-slice computed tomography
MTF	Modulation transfer function
ROI	Regions of interest
SD	Standard deviation
VOI	Volumes of interest

Introduction

The diagnosis of thin dento-maxillary and skull-base lesions requires high-definition bone imaging. Nowadays, X-ray computed tomography (CT) is the best available technical issue for studying tiny bone structures [1]. Over the last decade, manufacturers designed polyvalent multi-slice CT systems (MSCT) in order to reduce acquisition time, especially for cardiology, vascular, whole body and emergency imaging [2–4]. There are various technological approaches aimed at reducing acquisition times; for instance, increasing the number of rows for incremental acquisition or proposing dual-source imaging for dual helical acquisitions. Today, however, higher spatial resolution imaging is no longer a priority relative to the search for temporal resolution and radiation safety. Nevertheless, clinical applications concerning dento-maxillar and temporal bone regions [1] need higher resolution, as for instance when searching for a fistula tract in maxillar osteomyelitis, or superior canal dehiscence. More recently, a lower-cost specific architecture was designed for high-resolution skull-bone imaging using a flat panel detector (FPD) called cone-beam CT (CBCT) [5, 6]. We compare in this study three MSCT and one CBCT, from different manufacturers, that have been installed in our University Hospital since 2011. The purpose of this study is to evaluate linearity, geometric accuracy, homogeneity and spatial resolution for these apparatuses using a phantom adapted to small field of view (FOV) and high spatial resolution [7]. Each CT system was tested with protocols used in our clinical routine. To complete the quantitative analysis, a human dry skull was examined under the same conditions in order to allow for a visual qualitative evaluation.

Material and methods

CT systems and acquisition protocols

The MSCT instruments were: Somatom Flash CT (Siemens Medical, Erlangen, Germany); Discovery CT750HD (GE Healthcare, Milwaukee, USA); and Aquilion One 320 (Toshiba Medical Systems, Nasu, Japan). The CBCT was a Newtom 5G (QR s.r.l, Verona, Italy). All apparatuses were tested with modes that enable high-resolution bone imaging; they were all up-to-date in terms of hardware and software at the time of the study. We decided to work only with filtered back-projection (FPB) reconstruction algorithms, in order to make a comparison between all systems possible. Reconstructions were all done using the proprietary software. The characteristics of the CT systems and protocol acquisitions are

Table 1 System characteristics and protocol parameters

Manufacturer	Siemens	GE	Toshiba	Newtom DR srl
Model	Somatom Flash	Discovery CT 750 HD	Aquilion One 320	Newtom 5G
Detectors characteristics	Ultrafast ceramic 64 rows 912 detectors per row Row thickness 0.6 mm	Gemstone scintillator 128 rows 736 detectors per row Row thickness 0.625 mm	Solide state GOS 320 rows 896 detectors per row Row thickness 0.5 mm	Amorphous silicon Flat panel detector: 20×25 cm 790×960 Row thickness 0.26 mm
Focal spot dimension (mm ²)	0.6×0.7	1×0.7	0.9×0.8	0.3×0.3
Acquisition number	#1	#3	#5	#6
Institution protocol name	Sinus	Sinus	Bone 0.5 vol	15×12
kVp	140	140	120	110
Collimation in z direction (mm)	38.4	20	160	80
Acquisition type	Helical Pitch=0.8	Helical Pitch=0.53	Incremental	Incremental
Reconstructed voxel size (mm ³)	0.16×0.16×0.6	0.2×0.2×0.625	0.19×0.19×0.5	0.25×0.25×0.25
Rotation duration (360°)	1 s	1 s	1 s	18 s
Acquisition time (z=8 cm)	2.6 s	7.5 s	1 s	18 s
				22 s
				22 s

Fig. 1 Separate modular sections of the vmCT phantom acquired on a preclinical micro-CT system (eXplore CT 120, GE Healthcare, Waukesha, WI, USA). Resolution coil plate (**a**), linearity plate with iodine concentration listed in mg.ml^{-1} (**b**), Geometric accuracy plate with copper wire (**c**)

summarized in Table 1. Each acquisition has been numbered (from #1 to #8); this identification number will be used in the following for further reference. For MSCT, we used the proposed standard and high-resolution bone protocols. These two protocols were not available for the Toshiba system and we used the best-resolved acquisition protocol offered by the manufacturer.

Phantom

The vmCT (Shelley medical technologies, Toronto, Canada) is a phantom [6] consisting of separated modular sections, each designed to evaluate one particular aspect of image quality (resolution, geometric accuracy, linearity, uniformity and noise). The sections are held together inside an acrylic tube with an inner diameter of 63 mm and an outer diameter of 70 mm; the total length of the phantom is 54 mm (Fig. 1). These dimensions are adapted to the use of a small FOV especially for the CBCT acquisition #8 (Table 1).

Spatial resolution

The resolution coil plate of the phantom (Fig. 1) provides a visual qualitative measurement of the spatial resolution of the system. Embedded in the polycarbonate plastic plate of 4.8 mm thickness, there are four alternated aluminium and mylar sheets spiral coils, with layer thicknesses of 150, 200,

300 and 500 μm , corresponding to 3.3, 2.5, 1.67 and 1 line pairs per mm (lp.mm^{-1}), respectively. The standard deviation (SD) in four volumes of interest (VOI) of $1.5 \times 1.5 \times 1.5 \text{ mm}^3$ placed on each coil (top, bottom, left and right) were measured and corrected by the SD of the phantom uniform region. For the quantitative spatial resolution analysis, the modulation transfer function (MTF) values were calculated for each coil thickness using the average SD and the mean absolute difference CT values measured for aluminum and plastic [8, 9]. The reference value (MTF 100) was determined from the CT number for pure aluminum (CT_{Al}) and plastic ($\text{CT}_{\text{plastic}}$), according to formula (1) [9].

$$\text{MTF100} = \frac{|\text{CT}_{\text{Al}} - \text{CT}_{\text{plastic}}|}{2} \quad (1)$$

A subjective blinded visual analysis was conducted by ten trained CT users (radiologists and technologists) to qualitatively classify the different acquisitions. For each acquisition, a semiquantitative visual scale using three subscores evaluating spatial resolution, blurring and image artefacts, was applied (Table 2). The different acquisitions were ranked from these results. This blind analysis was performed using the free software Osirix (www.osirix-viewer.com).

Table 2 Phantom analysis scoring

Phantom analysis scale						
Spiral coil (500 μm)	non visible	0	lightly visible	1	clearly visible	2
Spiral coil (300 μm)	non visible	0	lightly visible	1	clearly visible	2
Spiral coil (200 μm)	non visible	0	lightly visible	1	clearly visible	2
Spiral coil (150 μm)	non visible	0	lightly visible	1	clearly visible	2
Blurring and artifacts	detrimental	0	troublesome	1	non visible	2

Table 3 Human skull analysis scoring

Dento-maxillar region analysis scale						
Dental and peridental	low quality	0	convenient	1	high quality	2
Alveolar bone	low quality	0	convenient	1	high quality	2
Blurring and artifacts	detrimental	0	troublesome	1	non visible	2
Temporal region analysis scale						
Petrous bone	low quality	0	convenient	1	high quality	2
Ear	low quality	0	convenient	1	high quality	2
Blurring and artifacts	detrimental	0	troublesome	1	non visible	2

Geometric accuracy

Four beads were placed 35 mm apart at the corners of an ideal square centred in the geometric accuracy section (Fig. 1), and one bead was placed at the centre of the square (24.75 mm from the other four). The distance between two neighbouring beads was measured between the beads centres on native slices. The differences between the known physical distances and the measured distances were expressed in % and the mean distance error of each acquisition was compared to its pixel size.

Linearity

The linearity of the system was determined by the linearity plate, which consisted of vials of increasing iodine concentrations of 0.9375, 1.875, 3.75, 7.5, 15, 30 mg.ml⁻¹ [6]. The mean signal intensity and standard deviation for each iodine concentration were measured in HU in a 3 × 3 mm³ VOI placed manually at the centre of each vial.

The relationship between signal intensity and iodine concentration was determined by linear regression analysis (Microsoft Excel, Redmond, Washington, USA).

Uniformity

Four peripheral regions of interest (ROI) and one central ROI of 3 × 3 × 1 mm³ were drawn on the uniformity section in order to quantitatively assess the variation in signal intensity from the centre to the periphery. The average differences in signal intensity between the peripheral and the central regions, as well as the average measured SD measured in the five ROIs were used to calculate the uniformity-to-noise ratio. The total measured noise can be considered as a quadrature summation of photon noise and system noise [10].

Human skull-bone analysis

A human dry skull was loaned from the collection of the Institute of Normal Anatomy of the Strasbourg Faculty of

Fig. 2 Modulation transfer function (MTF) of each protocol (#1 to #8; see Table 1) measured from the resolution coil plate

Fig. 3 Axial slices through the resolution plates for different protocols (#1 to #8; see Table 1). Window level: 700; window width: 4000

Medicine [11], in order to test different protocols on real bone architecture. A blind visual analysis was conducted by ten different trained CT users (radiologists and technologists) for classifying the different acquisitions. In order to be comparable over the different systems, a dental maxillary region and a

temporal region were examined to perform this visual qualitative analysis. For each acquisition, a semiquantitative visual scale, using three subscores that evaluated anatomical structures, blurring and image artefacts, was applied (Table 3). The scores were adjusted to ten and the different acquisitions were

Table 4 Phantom and human dry skull scoring results

Rank	Radiologic phantom		Dento-maxillary region		Temporal bone region		Mean	
	Score / 10 (<i>n</i> =10)	Acq. number	Score / 10 (<i>n</i> =10)	Acq. number	Score / 10 (<i>n</i> =10)	Acq. number	Score / 10	Acq. number
1	5.6	#4	8.8	#2	8.9	#2	7.6	#2
2	5.1	#2	7.7	#4	8.4	#4	7.2	#4
3	4.3	#8	6.9	#8	5.2	#3	5.0	#8
4	3.8	#1	5.0	#7	4.5	#1	4.2	#1
5	3.2	#3	4.3	#1	4.0	#5	3.7	#3
6	2.8	#7	3.3	#6	3.8	#8	3.5	#7
7	2.7	#5	2.7	#3	2.7	#7	2.8	#5
8	1.6	#6	1.7	#5	1.0	#6	2.0	#6

ranked from these results. This blind analysis was performed using the free software OsiriX (www.osirix-viewer.com).

Results

Spatial resolution

The in-plane MTF of each system was based on the analysis of the reconstructed images of the coils. For quantitative analysis, the MTF at 1, 1.67, 2, 50, 3.50 lp.mm⁻¹ of each acquisition are summarized in Fig. 2.

The visual qualitative analysis allowed the visualisation of the 500 µm coil for each acquisition protocol, the 300 µm coil was seen for the #2, #3, #4, #7 and #8 acquisitions, the 200 µm coil plate was only seen in the acquisition #8, and the 150 µm coil plate was never seen (Fig. 3). Scores and rankings issued from different acquisitions are given in Table 4. These rankings match the quantitatively observed results (Fig. 2).

Geometric accuracy

For each acquisition, the measured errors were smaller than the dimensions of the pixels (Table 4), meaning that the geometric accuracy is efficient for all acquisitions and systems.

Linearity

The linearity of the system was determined using the measured CT number in iodine solutions at various concentrations (Fig. 1). Linear correlation coefficients R^2 between measured CT number and iodine concentration are given in Table 5. The R^2 correlation coefficient value is comprised between 0.9809 and 0.9931, which shows that all systems are highly linear over the range of signal intensities encountered.

Uniformity

Uniformity of the system's signal response was measured on the polycarbonate uniformity section. The average difference in signal intensity values between peripheral and central regions is reported in Table 4. Contrary to CBCT protocols (#6

Table 5 Geometric accuracy, linearity and homogeneity results for each protocol (#1 to #7; see Table 1)

Acquisition number	#1	#2	#3	#4	#5	#6	#7	#8
Geometric accuracy								
Mean measurement error (mm)	0.08	0.04	0.07	0.10	0.11	0.02	0.05	0.011
Pixel size (mm)	0.16	0.16	0.20	0.17	0.19	0.25	0.15	0.075
Linearity								
R^2	0.9826	0.9896	0.9916	0.9884	0.9709	0.9931	0.9893	0.9924
Homogeneity								
Uniformity (SD) in CT value	5.5 (1.29)	11.5 (7)	1.75 (1.28)	2.5 (2.38)	6.53 (2.51)	28.5 (24.4)	35.03 (23.03)	33.52 (31.06)
Noise (SD) in CT value	43 (6.52)	27.2 (1.1)	22 (1.41)	33.2 (1.79)	25.83 (2.86)	16.18 (4.04)	22.56 (3.13)	34.84 (1.94)

and #8), we note that for MSCT protocols (#1 to #5) the average differences in signal intensity values between centre and periphery were all lower than the measured noise-signal intensity values. This means that the MSCT systems present a better uniformity than the CBCT system. Radial signal profiles of the uniformity section illustrate this point (Fig. 4).

Fig. 4 Radial signal profiles of the uniformity section for each protocols (#1 to #8; see Table 1)

Human skull-bone

Figure 5 shows a slice at the same position of the dental region obtained with each protocol (#1 to #8; see Table 1). Figure 6 presents axial slices through the right temporal bone (#1 to #8; see Table 1). Scores and ranking issued from these different

anatomical regions are given in Table 4. These results show that the trained users stated that acquisitions #2 and #4 stand out from the others by systematically taking the lead in ranking. The ranking of the other acquisitions varies

considerably from one expert to another, which precludes classifying them. In general, we observed that the CBCT acquisitions (#6 to #8) showed better scores in the dento-maxillary region than in the temporal one (Table 4).

Fig. 5 Axial slice at the same position of the dental region obtained with each protocol (#1 to #8; see Table 1) on the human dry skull. Window level: 700; window width: 4000

Discussion

CT models and methods

This study compares different high-resolution protocols between recent MSCT systems and one CBCT system. To our

knowledge, our work is the first study that compares apparatuses of the same generation with a dedicated, sized phantom. In fact, previous studies did not use objective comparison methods with small FOV adapted phantoms [12, 13], or did not compare devices of the same generation or high-resolution protocols [14].

Fig. 6 Axial slice at the same position of the temporal region obtained with each protocol (#1 to #8; see Table 1) on the human dry skull. Window level: 700; window width: 4000

Phantom and dry-skull analysis

Our results do not demonstrate the supremacy of CBCT for bone-skull imaging that has been related in several studies [12, 15]. Our results, coming both from a phantom as well as a dry skull, demonstrate that MSCT protocols #2 and #4 outperformed all the other MSCT and CBCT protocols (Figs. 2, 3, 5, 6, Table 5). However, the CBCT acquisition #8 provides interesting results. It is the only protocol showing the coil plate at 2.5 lp.mm^{-1} on the phantom. It was ranked at the third position on the dento-maxillary analysis. This confirms that CBCT systems can be considered as a good alternative to MSCT in dentistry.

These results strengthen the expectations raised in an earlier study [16], and illustrate that spatial resolution is a complex issue that depends on both detectors' characteristics (unitary elements number and size), geometrical aspects (focal spot size, beam collimation), and acquisition mode (incremental, helical, number of projections, rotation duration time), but also reconstruction parameters, like: kernel type, pixel reconstructed size and reconstruction algorithm. However, as in other studies [13, 14], the phantom used does not allow a measure of resolution in the z direction.

Image homogeneity and geometric accuracy

The CBCT offers a large z-direction acquisition field by using an FPD. This geometry presents the advantage of acquiring a whole volume in a single rotation. However, it is more sensitive to scattered radiation, which has an impact on image homogeneity [13, 17], as demonstrated in our experience (Fig. 4). In order to limit the consequences of the Compton effect, we decided to realize acquisitions on a small FOV-adapted phantom (Fig. 1) and on a dry skull (Figs. 5, 6). In terms of linearity and geometric accuracy, all systems present similar and robust results (Table 5).

Practical aspect

Our study is not exhaustive and is limited to the latest manufacturer models available at our university hospital. We also do not discuss other practical aspects, like the reconstruction speed and post-processing software. We note that, in our clinical practice, CBCT remains more accessible and available than MSCT systems. However, closer attention needs to be given to the management of patient motion artefacts in CBCT, due to rotation duration time (Table 1). For this reason, MSCT remains preferred in our practice in cases of bone trauma [1] and in the case of restless patients, especially in paediatrics. This additional point, which has not been evaluated in this work, is not in favour of CBCT. In this paper, we chose to limit our study to image properties and did not look at radiation

safety considerations. Different kinds of studies state that MSCT is more irradiative than CBCT [15, 18].

State of the art

In general, MSCT companies are well established in the field of CT imaging and have been so since the beginning, therefore having more experience than companies developing CBCT systems. Today, CBCT is still quite a new technology in the field of clinical bone imaging, and our results confirm that currently it does not surpass MSCT. However, CBCT technology is still evolving, and current studies are trying to solve some problems such as homogeneity, which could impact the spatial resolution and improve the visualization of soft tissue [13, 17, 19].

Acknowledgments Authors are grateful to Drs J.-L. Dietemann, A. Gangi, M. Ohana, C. Roy, F. Veillon, for access to the different devices. The authors gratefully acknowledge all technologists from the University Hospital of Strasbourg who participated in this study. Authors warmly thank General Electric for the loan of the phantom used in this study.

The scientific guarantor of this publication is P.Choquet. The authors of this manuscript declare no relationships with any companies, whose products or services may be related to the subject matter of the article. The authors state that this work has not received any funding. No complex statistical methods were necessary for this paper. Institutional Review Board approval was not required because no living subjects were used in this study.

References

- Desikan RS, James YC (2014) Imaging of temporal bone trauma. *Oper Tech Otolaryngol Head Neck Surg* 25:110–117
- Fuchs TO, Kachelriess M, Kalender WA (2000) System performance of multislice spiral computed tomography. *IEEE Eng Med Biol Mag* 19:63–70
- Mahesh M, Cody DD (2007) Physics of cardiac imaging with multiple-row detector CT. *Radiographics* 27:1495–1509
- Hassan A, Nazir SA, Alkaghi H (2011) Technical challenges of coronary CT angiography: today and tomorrow. *Eur J Radiol* 79: 161–171
- Miracle AC, Mukherji SK (2009) Conebeam CT of the head and neck, part 1: physical principles. *Am J Neuroradiol* 30:1088–1095
- Miracle AC, Mukherji SK (2009) Conebeam CT of the head and neck, part 2: clinical applications. *Am J Neuroradiol* 30:1285–1292
- Du LY, Umoh J, Nikolov NH, Pollmann SI, Lee TY, Holdsworth DW (2007) A quality assurance phantom for the performance evaluation of volumetric micro-CT systems. *Phys Med Biol* 52:7087–7108
- Droege RT (1983) A quality assurance protocol for CT scanners. *Radiology* 146:244–246
- Droege RT, Morin RL (1982) A practical method to measure the MTF of CT scanners. *Med Phys* 9:758–760
- Holdsworth DW, Drangova M, Fenster A (1993) A high-resolution XRII-based quantitative volume CT scanner. *Med Phys* 20:449–462
- Le Minor JM, Billmann F, Vetter JM, Ludes B (2009) Anatomie(s) & pathologie(s): Les collections morphologiques de la faculté de médecine de Strasbourg, 1st edn. ID L'Édition, Bernardswiller

12. Liang X, Jacobs R, Hassan B, Li L, Pauwels R, Corpas L, Souza PC, Martens W, Shahbazian M, Alonso A, Lambrechts I (2010) A comparative evaluation of Cone Beam Computed Tomography (CBCT) and Multi-Slice CT (MSCT) part I. On subjective image quality. *Eur J Radiol* 75:265–269
13. Carrino JA, Al Muhit A, Zbijewski W, Thawait GK, Stayman JW, Packard N, Senn R, Yang D, Foos DH, Yorkston J, Siewerdsen JH (2014) Dedicated cone-beam CT system for extremity imaging. *Radiology* 270:816–824
14. Pauwels R, Beinsberger J, Stamatakis H, Tsiklakis K, Walker A, Bosmans H, Bogaerts R, Jacobs R, Horner K (2012) Comparison of spatial and contrast resolution for cone-beam computed tomography scanners. *Oral Surg Oral Med Oral Pathol Oral Radiol* 114: 127–135
15. Hofmann E, Schmid M, Sedlmair M, Banckwitz R, Hirschfelder U, Lell M (2014) Comparative study of image quality and radiation dose of cone beam and low-dose multislice computed tomography—an in-vitro investigation. *Clin Oral Investig* 18:301–311
16. Watanabe H, Honda E, Tetsumura A, Kurabayashi T (2011) A comparative study for spatial resolution and subjective image characteristics of a multi-slice CT and a cone-beam CT for dental use. *Eur J Radiol* 77:397–402
17. Dong X, Petrongolo M, Niu T, Zhu L (2013) Low-dose and scatter-free cone-beam CT imaging using a stationary beam blocker in a single scan: phantom studies. *Comput Math Methods Med* 2013:637614
18. Loubele M, Bogaerts R, Van Dijk E, Pauwels R, Vanheusden S, Suetens P, Marchal G, Sanderink G, Jacobs R (2009) Comparison between effective radiation dose of CBCT and MSCT scanners for dentomaxillofacial applications. *Eur J Radiol* 71:461–468
19. Reitz I, Hesse BM, Nill S, Tücking T, Oelfke U (2009) Enhancement of image quality with a fast iterative scatter and beam hardening correction method for kV CBCT. *Z Med Phys* 19:158–172